


THE NEED TO IDENTIFY OCCUPATIONAL EXPOSURE TO LASER RADIATION IN GREECE

M. MAKROPOULOU,¹ C.J. HOURDAKIS,² A. SERAFETINIDES,¹ I. TSILIKAS,¹ I. SIANOUDIS,³
A. ACHTIPIS,⁴ C. KAPPAS,⁵ I. TSOUGOS,⁵ T.G. MARIS⁶ AND G. A. GOURZOULIDIS^{4,5}



¹National Technical University of Athens, Department of Physics, Athens 15780

²Licensing and Inspection Department, Greek Atomic Energy Commission

³Technological Educational Institute (TEI) of Athens, Dept of Optics & Optometry

⁴Hazardous Agents Department, OHS Directorate, Hellenic Ministry of Labor


⁵Medical Physics Department, Medical School, University of Thessaly

⁶Medical Physics Department, Medical School, University of Crete.

Corresponding author's e-mails: mmakro@central.ntua.gr, ggourz@yahoo.com

PURPOSE / INTRODUCTION

- Directive 2006/25/EC is a specific Occupational Health & Safety (OHS) one for Artificial Optical Radiation (AOR) - **UV, visible and IR.**
- Employs **limits** and **OHS measures** for **laser** and **non-coherent AOR.**
- Limited actions in Greece 10 years after the release.
- Ministry of Labor, NTUA and GAEC identify the field.



METHODS (1/2)



| Laser Safety Checklist | | Date: _____ |
|---|--|-------------|
| Institution/Dept: _____ | | Comments |
| System: _____ | | |
| The Laser system: _____ | | |
| Category: 1 1M 2 2M 3R 3B 4 | | |
| Operating Voltage: _____ Cooling: YES NO | | |
| Operating Environment: Temperature: _____ Humidity: _____ | | |
| Laser Beam / Output | | |
| Laser operation: CW - Pulsed Energy Output: _____ | | |
| Pulse Duration: _____ Δt Between Pulses: _____ | | |
| Frequency: _____ | | |
| Safety | | Comments |
| Danger Signs: Yes - No | | |
| Secured Area of the laser system: Yes - No | | |
| Protection Cover: Yes - No | | |
| Hazard Switch: Yes-No | | |
| Light Condition: VG - G - M - P - VP | | |
| Protective glasses: Yes - No | | |
| Protective for skin: Yes - No | | |
| Protective for eyes: Yes - No | | |

Lasers are widespread in:

- Hospitals
- Cosmetic applications
- Research
- Entertainment
- Industry

Assessment through:

- Evaluations of OHS regulations (**checklists** – laser experts)
- **Measurements** of appropriate optical quantities




METHODS (2/2)



- Safety issues:
 - Zone identification - signaling
 - Laser expert - instructions
 - **NOHA**
 - **PPE (OD>4)**
- Measuring equipment
 - **Energy meters**
 - Photometer
 - Thermohygrometer

- Measurements
 - **Main beam**
 - **Scattered beam**
 - **Limit verification (MPE)**
- Environmental conditions



RESULTS - COSMETICS

- **Nd:YAG** hair removal system (1064nm, pilot 650nm)
 - 10W mean power, 10ms pulse, 1Hz repetition rate, 4mm beam diam.
- Nominal output for **primary beam** 15J/cm²
 - Verification: 14.9 J/cm² (ophir novall)
 - 52,000 times over limit for eye per pulse
 - 8.5 times over skin limit per pulse

• Safety

- No signaling, no curtains, reflecting surfaces
- Glasses, no gloves



RESULTS – RESEARCH LAB (1/2)

- **Nd:YAG** research system
 - 6ns pulse, 1Hz repetition rate
- **Primary beam** nominal output
Verification: 120 mJ and 126 mJ
- **Safety**
 - Signaling, no curtains, reflecting surfaces, interlocks.
 - Glasses, no gloves
- Environmental Conditions in Lab: Humidity 55%, Temp. 27° C , 250 Lux .




RESULTS - RESEARCH LAB (2/2)

Specular
Reflection

Diffuse
Reflection




- **Scattered beam** measurements
 - 3 materials
 - Angles from catoptric reflection
 - Safety distance for worst case catoptric reflection (NOHA approach) \approx **2m**
 - Safety distance for diffuse reflection (NOHA approach) \approx **0.6m**



RESULTS - INDUSTRY

- Double head **Nd:YAG**, 1064nm, 500W mean power, 2.4J, 0.3ms pulse, 155Hz repetition rate.
- **Scattered beam:** great fluctuation – worst result: 25 times over eye limit, 0.1% of skin limit.
- **Safety**
 - Signaling, curtain, emergency buttons, no interlocks, PPE (OD>7), reflecting surfaces
- Environmental Conditions: Humidity 38%, Temp. 29° C , 200 Lux





DISCUSSION (1/2)



- Safety **measurements** & application of **Directive's limits** are possible
- Initial results reveal:
 - **Eye hazard is present** in installations (even from scattered beam)
 - **Safety procedures are not always followed**
- Objectives:
 - Creation of sample safety **checklists**
 - **Pilot safety procedures** for certain installations
 - Creation of sample risk assessment
 - **Laser safety expert** activation

A photograph of a safety checklist form. The form has several sections with checkboxes and lines for text. The sections include: 'Company Name', 'Date of Inspection', 'Inspector', 'Safety Measures', 'Temperature', 'Humidity', 'Laser Class', 'Operator', 'Risk Assessment', 'Signature', and 'Date'. There are also columns for 'Current' and 'Comments'.

DISCUSSION (2/2)

- Objectives (continued):
 - Medical safety and QA protocols
 - Legislative upgrade of laser safety
- Lack of overall data of laser installations
 - Pilot results for sample ones



REFERENCES

- Directive 2006/25/EC: Artificial Optical Radiation, OJ L 2006; 114/38.
- Non-binding guide to good practice for implementing Dir 2006/25/EC, ISBN 978-92-79-16046-2.
- MEDICAL LASERS: QUALITY CONTROL, SAFETY STANDARDS, AND REGULATIONS, AAPM REPORT NO. 73, 2001.
- Determining hazard distances from non-Gaussian lasers, Wesley J. Marshall, APPLIED OPTICS / Vol. 30, No. 6 / 20.2.1991.
- Laser safety Information Bulletin, Laser Institute of America ■ (407) 380-1553 ■ www.laserinstitute.org